



| IHE Integration Statement | | Date: Sep. 6 th , 2017 |
|---|---|--|
| Vendor | Product Name | SW Build |
| ESAOTE S.p.A. | MyLab 9 * * with DICOM [®] option | F07XXXX |
| These products are intended to implement all transactions required in the IHE Technical Framework to support the IHE Integration Profiles, Actors and Options listed below: | | |
| Integration Profiles Implemented | Actors Implemented | Options Implemented |
| Scheduled Workflow | Acquisition Modality | Broad Worklist Query |
| Patient Information Reconciliation | Acquisition Modality | none |
| Portable Data for Imaging | Portable Media Creator | none |
| Evidence Documents | Acquisition Modality | none |
| Echocardiography Workflow | Acquisition Modality | Broad Worklist Query Stress Echo |
| Cardiology Evidence Documents | Acquisition Modality | Echo Evidence |
| Consistent Time | Time Client | none |
| Internet address for vendor's IHE information: http://www.esaote.com | | |
| Vendor's e-mail address for information about IHE: ihe@esaote.com | | |
| Links to Standards Conformance Statements for the Implementation | | |
| DICOM[®] | http://www.esaote.com/dicom.htm | |
| Links to general information on IHE | | |
| In North America: www.ihe.net | In Europe: www.ihe-europe.net | In Japan: www.ihe-j.org |

ESAOTE S.p.A.
via Melen 77
16152 Genova (Italy)

© Copyright ESAOTE S.p.A. 2017. All rights reserved.

* DICOM is the registered trademark of the National Electrical Manufacturers Association for its standards publications relating to digital communications of medical information.